

**ANALYSENZERTIFIKAT**  
CERTIFICATE OF ANALYSIS

<b>Bayol F</b> reinst	<b>Kat.Nr./Cat.No. :</b>	<b>14500</b>
research grade	<b>Lot/Contr.No. :</b>	<b>130647</b>

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
<b>Aussehen</b> appearance		klare, farblose, viskose Flüssigkeit clear, colorless, viscous liquid	entspricht corresponds
<b>A 1 cm/ 260 nm</b>	10 % in Hexan	max 0.1	0.046
<b>A 1 cm/ 280 nm</b>	10 % in Hexan	max. 0.1	0.071
<b>Brechungsindex</b> refractive index	$n_D^{20}$	1.468 – 1.470	1.4690
<b>Mindesthaltbarkeit</b> minimum shelf life			05/ 2018
<b>Lagerung (°C)</b> storage			+15 bis/ to +30

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.**  
**Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
We do not guarantee that the product can be used for a special application.  
This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date: 31.05.2013**

Dipl.-Ing. (FH) Bernhard Göckel

Patricia Steinmann

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
This report has been computer generated and does not contain a signature.