

**A N A L Y S E N Z E R T I F I K A T**  
CERTIFICATE OF ANALYSIS

<b>Bayol F</b> reinst research grade	<b>Kat.Nr./Cat.No. :</b> 14500 <b>Lot/Contr.No. :</b> 200860
---	---

<b>Parameter</b> parameter	<b>Methode</b> method	<b>Spezifikation</b> specification	<b>Ergebnis</b> result
<b>Aussehen</b> appearance		klare, farblose, viskose Flüssigkeit clear, colourless, viscous liquid	entspricht corresponds
<b>Identität</b> identity	IR	entspricht corresponds	entspricht corresponds
<b>A 1 cm/ 260 nm</b>	10% in Hexan	max 0.1	0.024
<b>A 1 cm/ 280 nm</b>	10% in Hexan	max. 0.1	0.025
<b>Mindesthaltbarkeit</b> minimum shelf life			20.08.2025
<b>Lagerung (°C)</b> storage			+15 bis/ to +30

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.**  
**Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
We do not guarantee that the product can be used for a special application.  
This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date:** 25.08.2020

Christian Monsler

Daniela Lux-Helmstetter

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
This report has been computer generated and does not contain a signature.