

A N A L Y S E N Z E R T I F I K A T
 CERTIFICATE OF ANALYSIS

Bayol F reinst	Kat.Nr./Cat.No. :	14500
research grade	Lot/Contr.No. :	210975

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Aussehen appearance		klare, farblose, viskose Flüssigkeit clear, colourless, viscous liquid	entspricht corresponds
Identität identity	IR	entspricht corresponds	entspricht corresponds
A 1 cm/ 260 nm	10% in Hexan	max 0.1	0.025
A 1 cm/ 280 nm	10% in Hexan	max. 0.1	0.034
Mindesthaltbarkeit minimum shelf life			16.08.2026
Lagerung (°C) storage			+15 bis/ to +30

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
 We do not guarantee that the product can be used for a special application.
 This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 25.08.2021

Christian Monsler

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
 This report has been computer generated and does not contain a signature.