

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

| | |
|--|--------------------------------|
| Trichloressigsäure, 20 % Lösung | Kat.Nr./Cat.No. : 36913 |
| Trichloroacetic acid, 20 % solution | Lot/Contr.No. : P240152 |

| Parameter parameter | Spezifikation specification | Ergebnis result |
|---|---|---------------------------|
| Aussehen appearance | klare, farblose Lösung clear, colorless solution | entspricht corresponds |
| Gehalt TCA (g/l) assay TCA | 200.0 | entspricht corresponds |
| pH | < 0.5 | 0.33 |
| Mindesthaltbarkeit minimum shelf life | | 30.04.2026 |
| Lagerung (°C) storage | | +15 bis/ to +30 |

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control**Druckdatum/ printing date: 02.05.2024**

Dr. Judith Koch

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.