

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

Wasser DEPC behandelt, steril	Kat.Nr./Cat.No. :	39798
Water DEPC treated, sterile molecular biology grade	Lot/Contr.No. :	P160076

Parameter parameter	Spezifikation specification	Ergebnis result
Summenformel formula	H ₂ O	
Molekulargewicht molecular weight	18.0	
Aussehen appearance	klare, farblose Flüssigkeit clear, colorless liquid	entspricht corresponds
Wasser behandelt mit (% DEPC) water treatment	min. 0.1	min. 0.1
Leitfähigkeit (µS/ cm) conductivity	max. 10	2
Steriltest sterility	kein Wachstum no growth	entspricht corresponds
DNase/ RNase Aktivität DNase/ RNase activity	nicht nachweisbar none detected	entspricht corresponds
Mindesthaltbarkeit minimum shelf life		17.02.2017
Lagerung (°C) storage		+15 bis/ to +30

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 25.02.2016

Dipl.-Ing. (FH) Bernhard Göckel

Christian Monsler

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.