

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

TAE Puffer (40x), 1.2 M	Kat.Nr./Cat.No. : 42548
TAE buffer (40x)	Lot/Contr.No. : P150320

Parameter parameter	Spezifikation specification	Ergebnis result
Aussehen appearance	klare, farblose Lösung clear, colorless solution	entspricht corresponds
pH	8.3 – 8.7	8.4
Zusammensetzung composition		
TRIS-Acetat	1.2 M	entspricht corresponds
EDTA-Na ₂ ·2H ₂ O	0.04 M	entspricht corresponds
Mindesthaltbarkeit minimum shelf life		03/ 2017
Lagerung (°C) storage		+15 bis/ to +30

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 10.09.2015

Dipl.-Ing. (FH) Bernhard Göckel

Patricia Milford

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.