

A N A L Y S E N Z E R T I F I K A T
CERTIFICATE OF ANALYSIS

| | |
|--------------------------------|--------------------------------|
| TAE Puffer (40x), 1.2 M | Kat.Nr./Cat.No. : 42548 |
| TAE buffer (40x) | Lot/Contr.No. : P190195 |

| Parameter parameter | Spezifikation specification | Ergebnis result |
|---|---|---------------------------|
| Aussehen appearance | klare, farblose Lösung clear, colorless solution | entspricht corresponds |
| pH | 8.3 – 8.7 | 8.5 |
| Zusammensetzung composition | | |
| TRIS-Acetat | 1.2 M | entspricht corresponds |
| EDTA-Na ₂ ·2H ₂ O | 0.04 M | entspricht corresponds |
| Mindesthaltbarkeit minimum shelf life | | 23.11.2020 |
| Lagerung (°C) storage | | +15 bis/ to +30 |

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 24.05.2019

Christian Monsler

Patricia Milford

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.