

A N A L Y S E N Z E R T I F I K A T
CERTIFICATE OF ANALYSIS

TAE Puffer (40x), 1.2 M	Kat.Nr./Cat.No. : 42548
TAE buffer (40x)	Lot/Contr.No. : P210067

Parameter parameter	Spezifikation specification	Ergebnis result
Aussehen appearance	klare, farblose Lösung clear, colorless solution	entspricht corresponds
pH	8.3 – 8.7	8.5
Zusammensetzung composition		
TRIS-Acetat	1.2 M	entspricht corresponds
EDTA-Na ₂ ·2H ₂ O	0.04 M	entspricht corresponds
Mindesthaltbarkeit minimum shelf life		25.08.2022
Lagerung (°C) storage		+15 bis/ to +30

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 25.02.2021

Christian Monsler

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.