

ANALYSENZERTIFIKAT
 CERTIFICATE OF ANALYSIS

| | |
|---|--------------------------------|
| Sprüh-Magermilchpulver zum Blotten | Kat.Nr./Cat.No. : 42590 |
| Skim milk powder for blotting | Lot/Contr.No. : 170102 |

| Parameter parameter | Methode method | Spezifikation specification | Ergebnis result |
|---|--------------------------|---|---------------------------|
| Aussehen appearance | | weißes bis gelbliches Pulver white to yellowish powder | entspricht corresponds |
| Protein (%) protein | | 32.0 – 40.0 | 34.6 |
| Fett (%) fat | | max. 1,25 | 0.5 |
| Lactosemonohydrat (%) Lactose monohydrate | | 45.0 – 56.0 | entspricht corresponds |
| pH | 10 % in H ₂ O | 6.4 – 6.7 | 6.7 |
| Asche (%) ash | | max. 8.5 | 7.5 |
| Wassergehalt (%) water content | | max. 4.0 | 3.9 |
| Mindesthaltbarkeit minimum shelf life | | | 27.12.2018 |
| Lagerung (°C) storage | | | +15 bis/ to +30 |

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
 We do not guarantee that the product can be used for a special application.
 This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 13.02.2017

Dipl.-Ing. (FH) Bernhard Göckel

Christian Monsler

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
 This report has been computer generated and does not contain a signature.