

ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS



**SERVAGel(TM) TG PRiME(TM) 4 - 12 %**      **Kat.Nr./Cat.No. : 43273**  
**10 cm x 10 cm, 1 mm, 12 sample wells**      **Lot/Contr.No. : V24058-1**

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result	
<b>Beschreibung</b>	Gebrauchsfertige Polyacrylamid-Bisacrylamid Gele in einer Einweg-Plastikkassette zur Verwendung in der Gelelektrophorese von Proteinen. Format: 10 x 10 cm, Dicke: 1 mm, 12 Probetaschen			
description	Ready-to-use polyacrylamide-bisacrylamide gels in a disposable plastic cassette for use in gel electrophoresis of proteins. Format: 10 x10 cm, Thickness: 1 mm, 12 sample wells			
<b>Testbedingungen</b>	Elektrophorese der SERVA Protein-Testmischung 6 für SDS PAGE unter Standardbedingungen bis Lauffront Gelende erreicht.			
testing conditions	Electrophoresis of SERVA Protein Test Mixture 6 for SDS PAGE under standard conditions until running front has reached end of gel.			
<b>Proteintrennung</b>	Albumin (BSA)	67 kDa	entspricht/conform	entspricht/conform
protein separation	Carbonic anhydrase	29 kDa	entspricht/conform	entspricht/conform
	Trypsin inhibitor	21 kDa	entspricht/conform	entspricht/conform
<b>Aussehen</b>	Die Proteintrennung zeigt scharfe und gerade Banden über den gesamten Trennbereich des Gels.			
appearance	Protein separation shows sharp and straight bands over the whole separation range of the gel.			
<b>Mindesthaltbarkeit</b>			14.06.2025	
minimum shelf life				
<b>Lagerung (°C)</b>			+2 bis/ to +8	
storage				

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.**  
**Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
We do not guarantee that the product can be used for a special application.  
This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date: 16.05.2024**

Dr. Judith Koch

Daniela Lux-Helmstetter

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
This report has been computer generated and does not contain a signature.