

**ANALYSENZERTIFIKAT**  
CERTIFICATE OF ANALYSIS

<b>Gentamycin Lösung 50mg/ml, sterilfiltriert</b>	<b>Kat.Nr./Cat.No. : 47991</b>
Gentamycin solution 50mg/ml, sterile filtered	<b>Lot/Contr.No. : P110439</b>

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
<b>Aussehen</b> appearance		klare, leicht gelbliche Lösung clear, slightly yellowish solution	entspricht corresponds
<b>Konzentration (mg/ ml)</b> concentration	in H <sub>2</sub> O	80 mg Gentamycinsulfat entspricht 50 mg Gentamycin 80 mg Gentamycin sulfate corresponds 50 mg Gentamycin	
<b>pH</b>		4.0 – 6.0	4.7
<b>Osmolarität (mOsmol/ l)</b> osmolarity		ca. 100	133
<b>Sterilität</b> sterility		negativ negative	entspricht corresponds
<b>Mindesthaltbarkeit</b> minimum shelf life			04/ 2016
<b>Lagerung (°C)</b> storage			-15 bis/ to -25

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.**  
**Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
We do not guarantee that the product can be used for a special application.  
This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH** **Druckdatum/ printing date: 19.02.2015**  
**Qualitätskontrolle/ Quality Control**

Dipl.-Ing. (FH) Bernhard Göckel Daniela Lux-Helmstetter

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
This report has been computer generated and does not contain a signature.