

## PRODUCT INFORMATION

**ReadyLyzer 0.8 MWCO 1 kDa**

**Cat. No.: 44622**

### PRODUCT DESCRIPTION

ReadyLyzer are ready-to-use devices for quick and efficient dialysis consisting of a low-binding plastic tube with an ultra-pure dialysis membrane pre-installed and screw cap.

---

**Application** Desalting, buffer exchange, removal of small molecular impurities, sample concentration

---

### Dialysis Procedure

- Fill the ReadyLyzer with 800 µl dH<sub>2</sub>O, incubate it at least for 5 min and empty the tube.

---

**IMPORTANT:** Make sure that there is no dH<sub>2</sub>O leakage.

Absorption of water by the dry membrane will decrease the water level.

---

- Pipet sample (50 – 800 µl) into the ReadyLyzer tube and close the screw cap. Pipet small sample volumes (e.g. 50 µl) close to the inner membrane.
- Place the ReadyLyzer in the supplied floating rack and transfer it into a beaker (on a magnetic stirrer) containing a large volume of the desired buffer (100- to 1000-fold of the sample volume).
- Adjust the speed of the stir bar.
- Allow at least 30 min equilibration for each 100 µl of sample.  
Low-molecular salts and buffers (Tris-HCl, NaCl) equilibrate within 3 hours.  
Equilibration times for viscous samples will be longer.

---

**IMPORTANT:** Optimal equilibration times for the dialysis must be determined by the user.

---

- Change dialysis buffer as necessary, at least 2-3 times
  - After dialysis remove sample carefully using a pipette and transfer the sample into a clean tube.
- 

### Concentration Procedure

- Pipet sample into the ReadyLyzer or use already dialyzed sample, place it in bench top micro tube rack and let sample evaporate e.g. by using a fan to speed up the process.
- Check sample frequently to prevent complete evaporation.

---

**IMPORTANT:** When evaporation water, small molecules (buffer salts, reducing agents etc.) will also be concentrated in the sample.

---

**Storage** Store at + 2 °C to + 8 °C

Ver. 09/23