

**ANALYSENZERTIFIKAT**  
CERTIFICATE OF ANALYSIS

<b>Acrylamide 4x cryst. p.A.</b> analytical grade	<b>Kat.Nr./Cat.No. :</b> 10674 <b>Lot/Contr.No. :</b> 200758
--	---

<b>Parameter</b> parameter	<b>Methode</b> method	<b>Spezifikation</b> specification	<b>Ergebnis</b> result
<b>Summenformel</b> formula		C <sub>3</sub> H <sub>5</sub> NO	
<b>Molekulargewicht</b> molecular weight		71.08	
<b>Aussehen</b> appearance		weiße Kristalle white crystals	entspricht corresponds
<b>Gehalt (%)</b> assay	HPLC	min. 99.0	100.0
<b>Acrylsäure (%)</b> acrylic acid	Titration	max. 0.002	< 0.01
<b>A 1 cm/ 290 nm</b>	5 % in H <sub>2</sub> O	max. 0.6	0.524
<b>Löslichkeit</b> solubility	5 % in H <sub>2</sub> O	klar und farblos clear and colorless	entspricht corresponds
<b>pH</b>	5 % in H <sub>2</sub> O	5.0 – 8.0	6.2
<b>Leitfähigkeit (µS/cm)</b> conductivity	40 % in H <sub>2</sub> O, 25 °C	max. 20	2
<b>Polymerisationszeit (min)</b> polymerizing time	15 % Gel, 3 mM APS/ TEMED / 25 °C	max. 30	13 – 18
<b>Mindesthaltbarkeit</b> minimum shelf life			03.02.2022
<b>Lagerung (°C)</b> storage			+2 bis/ to +8

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.**  
**Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
We do not guarantee that the product can be used for a special application.  
This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date:** 27.07.2020

Christian Monsler

Patricia Milford

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
This report has been computer generated and does not contain a signature.