

**ANALYSENZERTIFIKAT**  
CERTIFICATE OF ANALYSIS

<b>Ethanol unvergällt absolut p.A.</b>	<b>Kat.Nr./Cat.No. : 11093</b>
Ethanol undenaturated absolute analytical grade	<b>Lot/Contr.No. : 150109</b>

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
<b>Summenformel</b> formula		C <sub>2</sub> H <sub>6</sub> O	
<b>Molekulargewicht</b> molecular weight		46.07	
<b>Aussehen</b> appearance		klare, farblose Flüssigkeit clear, colorless liquid	entspricht corresponds
<b>Gehalt (%)</b> assay	GC	min. 99.7	100.0
<b>Wasser (%)</b> water	KF	max. 0.2	0.02
<b>Brechungsindex</b> refractive index	n <sub>D</sub> <sup>20</sup>	1.3600 – 1.3620	1.3613
<b>Identität</b> identity	IR	entspricht corresponds	entspricht corresponds
<b>Mindesthaltbarkeit</b> minimum shelf life			01/ 2020
<b>Lagerung (°C)</b> storage			+15 bis/ to +30

<p><b>Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.</b>  <b>Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.</b>          We do not guarantee that the product can be used for a special application.          This document does not release you from performing the standard control upon receipt of incoming goods.</p>
--

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date: 27.01.2015**

Dipl.-Ing. (FH) Bernhard Göckel

Patricia Milford

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
 This report has been computer generated and does not contain a signature.