

**A N A L Y S E N Z E R T I F I K A T**  
 CERTIFICATE OF ANALYSIS

<b>Ethanol unvergällt absolut p.A.</b>	<b>Kat.Nr./Cat.No. : 11093</b>
Ethanol undenaturated absolute analytical grade	<b>Lot/Contr.No. : 200099</b>

<b>Parameter</b> parameter	<b>Methode</b> method	<b>Spezifikation</b> specification	<b>Ergebnis</b> result
<b>Summenformel</b> formula		C <sub>2</sub> H <sub>6</sub> O	
<b>Molekulargewicht</b> molecular weight		46.07	
<b>Aussehen</b> appearance		klare, farblose Flüssigkeit clear, colorless liquid	entspricht corresponds
<b>Gehalt V/V (%)</b> assay		min. 99.7	100.0
<b>Wasser (%)</b> water	KF	max. 0.2	0.02
<b>Brechungsindex</b> refractive index	n <sub>D</sub> <sup>20</sup>	1.3600 – 1.3620	1.3613
<b>Mindesthaltbarkeit</b> minimum shelf life			28.01.2025
<b>Lagerung (°C)</b> storage			+15 bis/ to +30

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.**  
**Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
 We do not guarantee that the product can be used for a special application.  
 This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date: 30.01.2020**

Christian Monsler

Daniela Lux-Helmstetter

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
 This report has been computer generated and does not contain a signature.