

A N A L Y S E N Z E R T I F I K A T
 CERTIFICATE OF ANALYSIS

| | |
|---|--------------------------------|
| Ethanol unvergällt absolut p.A. | Kat.Nr./Cat.No. : 11093 |
| Ethanol undenaturated absolute analytical grade | Lot/Contr.No. : 210235 |

| Parameter parameter | Methode method | Spezifikation specification | Ergebnis result |
|---|------------------------------|--|---------------------------|
| Summenformel formula | | C ₂ H ₆ O | |
| Molekulargewicht molecular weight | | 46.07 | |
| Aussehen appearance | | klare, farblose Flüssigkeit clear, colorless liquid | entspricht corresponds |
| Gehalt V/V (%) assay | | min. 99.7 | 100.0 |
| Wasser (%) water | KF | max. 0.2 | 0.04 |
| Brechungsindex refractive index | n _D ²⁰ | 1.3600 – 1.3620 | 1.3613 |
| Mindesthaltbarkeit minimum shelf life | | | 19.02.2026 |
| Lagerung (°C) storage | | | +15 bis/ to +30 |

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
 We do not guarantee that the product can be used for a special application.
 This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 26.02.2021

Christian Monsler

Patricia Milford

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
 This report has been computer generated and does not contain a signature.