

**A N A L Y S E N Z E R T I F I K A T**  
 CERTIFICATE OF ANALYSIS

<b>Ethanol unvergällt 96 % p.A.</b>	<b>Kat.Nr./Cat.No. :</b>	<b>11094</b>
Ethanol undenaturated 96 % analytical grade	<b>Lot/Contr.No. :</b>	<b>200497</b>

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
<b>Summenformel</b> formula		C <sub>2</sub> H <sub>6</sub> O	
<b>Molekulargewicht</b> molecular weight		46.07	
<b>Aussehen</b> appearance		klare, farblose Flüssigkeit clear, colorless liquid	entspricht corresponds
<b>Wasser (%)</b> water	KF	4.0 – 6.0	5.7
<b>Brechungsindex</b> refractive index	n <sub>D</sub> / 20 °C	1.3625 – 1.3635	1.3632
<b>Identität</b> identity	IR	entspricht corresponds	entspricht corresponds
<b>Mindesthaltbarkeit</b> minimum shelf life			12.05.2025
<b>Lagerung (°C)</b> storage			+15 bis/ to +30

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.**  
**Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
 We do not guarantee that the product can be used for a special application.  
 This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date: 18.05.2020**

Christian Monsler

Patricia Milford

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
 This report has been computer generated and does not contain a signature.