

**ANALYSENZERTIFIKAT**  
CERTIFICATE OF ANALYSIS

Ethylenglycol p.A. Ethylene glycol analytical grade	Kat.Nr./Cat.No. : Lot/Contr.No. :	11285 160927
--	--------------------------------------	-----------------

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
<b>Summenformel</b> formula		C <sub>2</sub> H <sub>6</sub> O <sub>2</sub>	
<b>Molekulargewicht</b> molecular weight		62.07	
<b>Aussehen</b> appearance		klare, farblose Flüssigkeit clear, colorless liquid	entspricht corresponds
<b>Identität</b> identity	IR	entspricht corresponds	entspricht corresponds
<b>Gehalt (%)</b> assay	GC	min. 99.5	99.9
<b>Mindesthaltbarkeit</b> minimum shelf life			28.02.2019
<b>Lagerung (°C)</b> storage			+15 bis/ to +30

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.**  
**Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
We do not guarantee that the product can be used for a special application.  
This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date: 30.09.2016**

Dipl.-Ing. (FH) Bernhard Göckel

Christian Monsler

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
This report has been computer generated and does not contain a signature.