

CERTIFICATE OF ANALYSIS

Albumin bovine Fraction V, pH 7.0 standard grade, lyophil.	Cat.No. : 11930 Contr.No.: 13182
--	---

Parameter	Method	Specification	Result
Protein (%) (protein factor 6.22)	Nitrogen Analyzer dry weight basis	96 – 100	98
Purity (%) (albumin, 7% solution)	Cellulose Acetate Electrophoresis	98 – 100	100
pH	7% solution	6.8 - 7.2	6.9
Sodium (mg/g)	Flame Photometer	0 - 10.0	3.4
Chloride (mg/g)	Coulometric Titrator	0 - 3.0	0.5
Iron (µg/g)	Atomic Absorption	0 - 15.0	4.8
IgG	Radialimmunodiffusion	not detectable	none detected
Moisture (%)	Karl Fischer	0 - 5.0	< 5.0
Expiry date			02/ 2005
Storage (°C)			+2 to +8

At time of shipment, this product qualifies for export. It was derived from bovine blood collected at a USDA licensed establishment. The cattle received ante- and post-mortem health inspection at the abattoir by a U.S. Veterinary Services Inspector and they were apparently free from infectious and contagious diseases. All donor animals were sourced from the United States, a country in which Bovine Spongiform Encephalopathy is not known to exist. Blood collection records are incorporated into the manufacturing batch documents.

<p>Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert. Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle. We do not guarantee that the product can be used for a special application. This document does not release you from performing the standard control upon receipt of incoming goods.</p>
--

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 06.11.2015

Dr. H. Volger

B. Göckel

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.

This report has been computer generated and does not contain a signature.