

**ANALYSENZERTIFIKAT**  
 CERTIFICATE OF ANALYSIS

<b>Ammoniumpersulfat p.A.</b>	<b>Kat.Nr./Cat.No. :</b>	<b>13375</b>
Ammonium persulfate analytical grade	<b>Lot/Contr.No. :</b>	<b>211042</b>

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
<b>Summenformel</b> formula		(NH <sub>4</sub> ) <sub>2</sub> S <sub>2</sub> O <sub>8</sub>	
<b>Molekulargewicht</b> molecular weight		228.20	
<b>Aussehen</b> appearance		weiße Kristalle white crystals	entspricht corresponds
<b>Gehalt (%)</b> assay	Titration	min. 99.0	99.9
<b>pH</b>	5 % in H <sub>2</sub> O	3.2 – 3.9	3.7
<b>Identität</b> identity	IR	entspricht corresponds	entspricht corresponds
<b>Mindesthaltbarkeit</b> minimum shelf life			08.09.2026
<b>Lagerung (°C)</b> storage			+15 bis/ to +30

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.**  
**Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
 We do not guarantee that the product can be used for a special application.  
 This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**
**Druckdatum/ printing date: 13.09.2021**

Christian Monsler

Marina Konter

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
 This report has been computer generated and does not contain a signature.