

**A N A L Y S E N Z E R T I F I K A T**  
CERTIFICATE OF ANALYSIS

<b>Ammoniumsulfat p.A.</b>	<b>Kat.Nr./Cat.No. :</b>	<b>13378</b>
Ammonium sulfate analytical grade	<b>Lot/Contr.No. :</b>	<b>180413</b>

<b>Parameter</b> parameter	<b>Methode</b> method	<b>Spezifikation</b> specification	<b>Ergebnis</b> result
<b>Summenformel</b> formula		(NH <sub>4</sub> ) <sub>2</sub> SO <sub>4</sub>	
<b>Molekulargewicht</b> molecular weight		132.2	
<b>Aussehen</b> appearance		weiße Kristalle white crystals	entspricht corresponds
<b>Gehalt (%)</b> assay	Titration	min. 99.0	99.0
<b>pH</b>	10 % in H <sub>2</sub> O	4.5 – 5.5	5.1
<b>Identität</b> identity	IR	entspricht corresponds	entspricht corresponds
<b>Mindesthaltbarkeit</b> minimum shelf life			10.04.2023
<b>Lagerung (°C)</b> storage			+15 bis / to +30

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.**  
**Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
We do not guarantee that the product can be used for a special application.  
This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date: 17.04.2018**

Dipl.-Ing. (FH) Bernhard Göckel

Christian Monsler

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
This report has been computer generated and does not contain a signature.