

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

Borsäure	Kat.Nr./Cat.No. :	15166
Boric acid electrophoresis grade	Lot/Contr.No. :	140682

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Summenformel formula		H ₃ BO ₃	
Molekulargewicht molecular weight		61.83	
Aussehen appearance		weißes, kristallines Pulver white, crystalline powder	entspricht corresponds
Gehalt (%) assay	Titration	99.0 – 100.5	100.0
pH	Ph.Eur.	3.8 – 4.8	4.1
Sulfat (ppm) sulfate	Ph.Eur.	max. 450	5
Schwermetalle (ppm) heavy metals	als/ as Pb	max. 15	5
Identität identity	IR	entspricht corresponds	entspricht corresponds
Anwendung application	DNA-Elektrophorese DNA-electrophoresis	entspricht corresponds	entspricht corresponds
Mindesthaltbarkeit minimum shelf life			07/ 2018
Lagerung (°C) storage			+15 bis/ to +30

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH **Druckdatum/ printing date: 16.06.2014**
Qualitätskontrolle/ Quality Control

Dipl.-Ing. (FH) Bernhard Göckel Patricia Milford

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.