

**ANALYSENZERTIFIKAT**  
CERTIFICATE OF ANALYSIS

|   |                                |
|---|--------------------------------|
| <b>BCIP/NBT, gebrauchsfertiges Substrat für<br/>Immunhistochemie und Blotting</b> | <b>Kat.Nr./Cat.No. : 15246</b> |
| BCIP/NBT, ready-to-use substrate for<br>Immunohistochemistry and Blotting         | <b>Lot/Contr.No. : 140236</b>  |

| <b>Parameter</b><br>parameter                   | <b>Methode</b><br>method | <b>Spezifikation</b><br>specification   | <b>Ergebnis</b><br>result |
|---|--------------------------|---|---------------------------|
| <b>Aussehen</b><br>appearance                   |                          | farblose bis schwach<br>gelbliche Lösung<br>colourless to pale<br>yellow solution | entspricht<br>corresponds |
| <b>pH</b>                                       |                          | 9,7 – 9,9   | entspricht<br>corresponds |
| <b>Anwendung</b><br>application                 | Dot Blot                 | entspricht<br>corresponds   | entspricht<br>corresponds |
| <b>Mindesthaltbarkeit</b><br>minimum shelf life |                          |   | 02/ 2016                  |
| <b>Lagerung (°C)</b><br>storage                 |                          |   | +2 bis/ to +8             |

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.  
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
We do not guarantee that the product can be used for a special application.  
This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date: 20.02.2014**

Dipl.-Ing. (FH) Bernhard Göckel

Patricia Milford

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
This report has been computer generated and does not contain a signature.