

**ANALYSENZERTIFIKAT**  
CERTIFICATE OF ANALYSIS

**Celite 545 pract.**

**Kat.Nr./Cat.No. : 16391**

**Lot/Contr.No. : 130224**

| Parameter<br>parameter                          | Spezifikation<br>specification             | Ergebnis<br>result        |
|-------------------------------------------------|--------------------------------------------|---------------------------|
| <b>Aussehen</b><br>appearance                   | grau-weißes Pulver<br>grey to white powder | entspricht<br>corresponds |
| <b>Korngröße (%)</b><br>particle size           | < 36 µm                                    | 33                        |
|                                                 | 36 - 45 µm                                 | 8                         |
|                                                 | 45 - 80 µm                                 | 27                        |
|                                                 | > 80 µm                                    | 32                        |
| <b>Trocknungsverlust (%)</b><br>loss on drying  | max. 0,5                                   | 0.1                       |
| <b>Mindesthaltbarkeit</b><br>minimum shelf life |                                            | 02/ 2018                  |
| <b>Lagerung (°C)</b><br>storage                 |                                            | +15 bis/ to +30           |

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.**

**Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**

We do not guarantee that the product can be used for a special application.

This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date: 25.02.2013**

Dipl.-Ing. (FH) Bernhard Göckel

Patricia Steinmann

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**

This report has been computer generated and does not contain a signature.