

**A N A L Y S E N Z E R T I F I K A T**  
CERTIFICATE OF ANALYSIS

<b>DePeX</b>	<b>Kat.Nr./Cat.No. :</b>	<b>18243</b>
	<b>Lot/Contr.No. :</b>	<b>120029</b>

<b>Parameter</b> parameter	<b>Methode</b> method	<b>Spezifikation</b> specification	<b>Ergebnis</b> result
<b>Aussehen</b> appearance		farblose, viskose Flüssigkeit colorless, viscous liquid	entspricht corresponds
<b>Zusammensetzung</b> composition		neutrale Lösung von Polystyrol und Weichmachern in Xylol neutral solution of polystyrene and plasticizers in xylene	
<b>Brechungsindex</b> refractive index	$n_D^{20^\circ C}$	1.52 – 1.53	1.5268
<b>Mindesthaltbarkeit</b> minimum shelf life			12/ 2014
<b>Lagerung (°C)</b> storage			+15 bis/ to +30

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.**  
**Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
We do not guarantee that the product can be used for a special application.  
This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date: 12.01.2012**

Dipl.-Ing. (FH) Bernhard Göckel

Christian Monsler

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
This report has been computer generated and does not contain a signature.