

**A N A L Y S E N Z E R T I F I K A T**  
CERTIFICATE OF ANALYSIS

**D.E.R. 736** pract.

**Kat.Nr./Cat.No. : 18247**

**Lot/Contr.No. : 180546**

<b>Parameter</b> parameter	<b>Methode</b> method	<b>Spezifikation</b> specification	<b>Ergebnis</b> result
<b>Aussehen</b> appearance		klare farblose bis gelbliche, leicht viskose Flüssigkeit clear colorless to yellowish, slightly viscous liquid	entspricht corresponds
<b>Polymerisation</b>		entspricht corresponds	entspricht corresponds
<b>Dichte</b> density		1.129 – 1.150	1.131
<b>Mindesthaltbarkeit</b> minimum shelf life			14.05.2023
<b>Lagerung (°C)</b> storage			+15 bis/ to +30

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.**  
**Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
We do not guarantee that the product can be used for a special application.  
This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date: 15.05.2018**

Dipl.-Ing. (FH) Bernhard Göckel

Christian Monsler

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
This report has been computer generated and does not contain a signature.