

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

Dextran 4 techn. grade	Kat.Nr./Cat.No. : 18687
	Lot/Contr.No. : 121303

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Aussehen appearance		weißes bis schwach grauweißes Pulver white to slightly off-white powder	entspricht corresponds
Mittleres Molekulargewicht average molecular weight		3500 – 5500	3532
Trocknungsverlust (%) loss on drying		max. 10.0	4.4
pH	10 % in H ₂ O	4.5 – 7.0	4.7
Mindesthaltbarkeit minimum shelf life			10/ 2016
Lagerung (°C) storage			+15 bis / to +30

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 06.08.2015

Dipl.-Ing. (FH) Bernhard Göckel

Patricia Milford

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.