

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

Dextran 4 techn. grade	Kat.Nr./Cat.No. :	18687
	Lot/Contr.No. :	181176

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Aussehen appearance		weißes bis schwach grauweißes Pulver white to slightly off-white powder	entspricht corresponds
Mittleres Molekulargewicht average molecular weight		3500 – 7500	6514
Trocknungsverlust (%) loss on drying		max. 10.0	4.2
pH	10 % in H ₂ O	4.0 – 7.0	4.9
Mindesthaltbarkeit minimum shelf life			31.08.2021
Lagerung (°C) storage			+15 bis / to +30

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 12.11.2018

Christian Monsler

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.