

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

Dextran 100 techn. grade	Kat.Nr./Cat.No. : 18693
	Lot/Contr.No. : 170263

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Molekulargewicht molecular weight		100 000 – 200 000	
Aussehen appearance		weißes Pulver white powder	entspricht corresponds
Trocknungsverlust (%) loss on drying		max. 7.0	5.2
pH	6 % in H ₂ O	4.5 – 7.0	7.0
Asche (%) ash		max. 2.0	0.5
Mindesthaltbarkeit minimum shelf life			30.09.2021
Lagerung (°C) storage			+15 bis/ to +30

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 15.07.2019

Christian Monsler

Patricia Milford

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.