

A N A L Y S E N Z E R T I F I K A T
 CERTIFICATE OF ANALYSIS

Dextran 100 , technische Qualität	Kat.Nr./Cat.No. :	18693
Dextran 100, technical grade	Lot/Contr.No. :	200962

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Molekulargewicht molecular weight		90 000 – 110 000	95 000
Aussehen appearance		weißes Pulver white powder	entspricht corresponds
Trocknungsverlust (%) loss on drying		max. 9.0	2.1
pH	6% in H ₂ O	4.5 - 7.0	5.8
Asche (%) ash		max. 2.0	0.18
Mindesthaltbarkeit minimum shelf life			15.09.2025
Lagerung (°C) storage			+15 bis/ to +30

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
 We do not guarantee that the product can be used for a special application.
 This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 29.09.2020

Christian Monsler

Patricia Milford

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
 This report has been computer generated and does not contain a signature.