

**ANALYSENZERTIFIKAT**  
CERTIFICATE OF ANALYSIS

<b>Diethylpyrocarbonat</b> reinst	<b>Kat.Nr./Cat.No. :</b>	<b>18835</b>
Diethyl pyrocarbonate research grade	<b>Lot/Contr.No. :</b>	<b>120647</b>

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
<b>Summenformel</b> formula		C <sub>6</sub> H <sub>10</sub> O <sub>5</sub>	
<b>Molekulargewicht</b> molecular weight		162.14	
<b>Aussehen</b> appearance		klare, farblose Flüssigkeit clear, colorless liquid	entspricht corresponds
<b>Gehalt (%)</b> assay	GC	min. 96.0	99.6
<b>Brechungsindex</b> refractive index	n <sub>D</sub> <sup>20</sup>	1.3960 – 1.4000	1.3970
<b>Identität</b> identity	IR	entspricht corresponds	entspricht corresponds
<b>Mindesthaltbarkeit</b> minimum shelf life			11/ 2013
<b>Lagerung (°C)</b> storage			+2 bis/ to +8

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.**  
**Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
We do not guarantee that the product can be used for a special application.  
This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date:** 26.06.2012

Dipl.-Ing. (FH) Bernhard Göckel

Daniela Lux-Helmstetter

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**

This report has been computer generated and does not contain a signature.