

**A N A L Y S E N Z E R T I F I K A T**  
 CERTIFICATE OF ANALYSIS

<b>20% SDS Lösung</b>	<b>Kat.Nr./Cat.No. : 20767</b>
20% SDS Solution	<b>Lot/Contr.No. : P180353</b>

<b>Parameter</b> parameter	<b>Spezifikation</b> specification	<b>Ergebnis</b> result
<b>Aussehen</b> appearance	klare, farblose bis leicht gelbliche Lösung clear, colorless to slightly yellowish solution	entspricht corresponds
<b>Gehalt SDS (g/l)</b> assay SDS	200	entspricht corresponds
<b>Farbe</b> color	max. 30 Hazen	5
<b>A 1cm/ 260 nm</b>	max. 0.4	0.029
<b>A 1cm/ 280 nm</b>	max. 0.15	0.024
<b>Mindesthaltbarkeit</b> minimum shelf life		25.10.2019
<b>Lagerung (°C)</b> storage		+15 bis/ to +30

Bei Lagerung unter +18°C kann das SDS auskristallisieren.  
 Die Kristalle lösen sich beim Erwärmen des Puffers wieder.

Storage below +18°C may cause crystallization of SDS.  
 Crystals can be solved by warming up the buffer solution.

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.**  
**Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
 We do not guarantee that the product can be used for a special application.  
 This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date: 26.10.2018**

Christian Monsler

Daniela Lux-Helmstetter

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
 This report has been computer generated and does not contain a signature.