

**ANALYSENZERTIFIKAT**  
CERTIFICATE OF ANALYSIS

<b>20 % SDS Lösung</b> 20 % SDS Solution electrophoresis grade	<b>Kat.Nr./Cat.No. :</b> <b>20768</b>
	<b>Lot/Contr.No. :</b> <b>P250384</b>

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
<b>Aussehen</b> appearance		klare, farblose bis leicht gelbliche Lösung clear, colorless to slightly yellowish solution	entspricht corresponds
<b>Gehalt SDS (g/l)</b> assay SDS		200	entspricht corresponds
<b>Farbe</b> color		max. 30 Hazen	10
<b>A 1cm/ 260 nm</b>		max. 0.4	0.078
<b>A 1cm/ 280 nm</b>		max. 0.15	0.035
<b>Anwendung</b> application	Elektrophorese electrophoresis	entspricht corresponds	entspricht corresponds
<b>Mindesthaltbarkeit</b> minimum shelf life			21.02.2027
<b>Lagerung (°C)</b> storage			+15 bis/ to +30

Bei Lagerung unter +18°C kann das SDS auskristallisieren.  
Die Kristalle lösen sich beim Erwärmen des Puffers wieder.

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.

Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.

We do not guarantee that the product can be used for a special application.

This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 27.11.2025

Dr. Judith Koch

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.

This report has been computer generated and does not contain a signature.