

ANALYSENZERTIFIKAT

CERTIFICATE OF ANALYSIS

Ethidiumbromid Lösung 1 % in Wasser	Kat.Nr./Cat.No. : 21251
Ethidiumbromide solution 1 % in water	Lot/Contr.No. : P210004

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Aussehen appearance		rote Lösung red solution	entspricht corresponds
λ_{\max} (nm)	0.001 % in H ₂ O	$\lambda_{\max 1} : 287 \pm 3$ $\lambda_{\max 2} : 483 \pm 4$	285 480
A 1 cm/ $\lambda_{\max 1}$	0.001 % in H ₂ O	min. 0.9	1.243
A 1 cm/ $\lambda_{\max 2}$	0.001 % in H ₂ O	min. 0.1	0.129
Mindesthaltbarkeit minimum shelf life			13.01.2025
Lagerung (°C) storage			+2 bis/ to +8

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
 We do not guarantee that the product can be used for a special application.
 This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 15.01.2021

Christian Monsler

Patricia Milford

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
 This report has been computer generated and does not contain a signature.