

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

Ficoll® 400	Kat.Nr./Cat.No. : 21373
	Lot/Contr.No. : 160172

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Molekulargewicht molecular weight		400 000 ± 100 000	321 030
Aussehen appearance		weißes Pulver white powder	entspricht corresponds
Trocknungsverlust (%) loss on drying		max. 5.0	3
Gehalt an NaCl (%) content of NaCl		max. 1.0	≤ 1.0
Mindesthaltbarkeit minimum shelf life			31.03.2020
Lagerung (°C) storage			+15 bis/ to +30

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 16.02.2016

Dipl.-Ing. (FH) Bernhard Göckel

Patricia Milford

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.