

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

Fluoromount für die Mikroskopie	Kat.Nr./Cat.No. :	21645
Fluoromount for microscopy	Lot/Contr.No. :	250160

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Aussehen appearance		klare, farblose, viskose Flüssigkeit clear, colorless, viscous liquid	entspricht corresponds
Dichte (g/ml) specific density		ca. 0.87	0.869
Brechungsindex refractive index	n_D^{20}	1.452 – 1.455	1.453
Mindesthaltbarkeit minimum shelf life			31.12.2026
Lagerung (°C) storage			+2 bis/ to +8

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 05.03.2025

Dr. Judith Koch

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.