

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

Gelrite®	Kat.Nr./Cat.No. :	22168
	Lot/Contr.No. :	120400

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Aussehen appearance		weißes Pulver white powder	entspricht corresponds
Gelstärke (g/cm²) gel strength		≥ 700	961
Trocknungsverlust (%) loss on drying		6.0 – 14.0	8.2
pH	0.5 % Lösung/ solution	6.0 – 8.0	6.4
Mindesthaltbarkeit minimum shelf life			09/ 2013
Lagerung (°C) storage			+15 bis/ to +30

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 28.03.2012

Dipl.-Ing. (FH) Bernhard Göckel

Christian Monsler

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.