

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

Gelrite®	Kat.Nr./Cat.No. : 22168
	Lot/Contr.No. : 190398

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Aussehen appearance		weißes Pulver white powder	entspricht corresponds
Gelstärke (g/ cm²) gel strength		min. 700	1489
Trocknungsverlust (%) loss on drying		max. 15	9.7
Transmission (%) transmittance	0.5% Gel (20 °C) 497 nm	min. 70	83.8
Mindesthaltbarkeit minimum shelf life			23.04.2022
Lagerung (°C) storage			+15 bis/ to +30

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 29.04.2019

Christian Monsler

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.