

CERTIFICATE OF ANALYSIS / ORIGIN

γ-Globulin bovine
lyophil.

Cat.No.: 22550
Contr.No.: 180972

Parameter	Method	Specification	Result
Protein (%) (protein factor 6.25)	Nitrogen Analyzer dry weight basis	96 – 100	99
Purity (%) (γ-globulins, 7 % solution)	Cellulose Acetate Electrophoresis	96 – 100	100
pH	7 % solution	6.8 – 7.2	7.0
Sodium (mg/g)	Flame Photometer	0 – 10.0	8.4
Chloride (mg/g)	Coulometric Titrator	0 – 24.0	11.8
Moisture (%)	KF	0 – 5.0	< 5.0
Minimum shelf life			31.08.2021
Storage (°C)			+2 to +8

The product has been subjected to a pH 5 or lower for a period of at least 3 hours.

It is certified that the product is derived from bovine blood collected from US sourced cattle slaughtered at a USDA licensed establishment located in the USA. The bovine plasma/serum was derived from US sourced animals under 30 months of age that were not stunned using a penetrating device that injects air into the cranial cavity. During the collection process appropriate precautions were taken not to contaminate the serum/plasma with specified risk materials (SRM) as defined by the USDA. The material does not contain, nor is derived from SRM as defined in REGULATION (EC) 999/2001. All cattle received ante- and post mortem health inspection under a veterinarian's supervision at the abattoir and were apparently free from infectious and contagious diseases and injurious parasites. Ruminant materials used in the manufacture of this product are not originated from BSE related herds. The record of each raw material collection is incorporated into the manufacturing records and veterinary certificates are maintained on file at the manufacturing site. At the time of manufacture of this lot of material, the US is classified as an OIE negligible BSE risk country in accordance with Chapter 11.5 of the Terrestrial Code and is free from rinderpest, foot-and-mouth disease, and contagious bovine pleuropneumonia.

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.

Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.

We do not guarantee that the product can be used for a special application.

This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 13.09.2018

Christian Monsler

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.

This report has been computer generated and does not contain a signature.