

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

Glutaraldehyd 25% in Wasser für Elektronenmikroskopie	Kat.Nr./Cat.No. :	23114
hoch rein	Lot/Contr.No. :	P140131
Glutaraldehyde 25% in water for electron microscopy		
high purity		

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Summenformel formula		C ₅ H ₈ O ₂	
Molekulargewicht molecular weight		100.13	
Aussehen appearance		klare, farblose Lösung clear, colorless solution	entspricht corresponds
Gehalt (%) assay		25.0 ± 1.0	24.2
pH		3.0 – 5.5	3.4
Purification index	A 235/ A 280, 1 % in H ₂ O	max. 0.5	0.130
Mindesthaltbarkeit minimum shelf life			03/ 2016
Lagerung (°C) storage			+2 bis/ to +8

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
 We do not guarantee that the product can be used for a special application.
 This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 12.03.2014

Dipl.-Ing. (FH) Bernhard Göckel

Patricia Milford

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
 This report has been computer generated and does not contain a signature.