

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

Glutaraldehyd 50 % in Wasser für Elektronenmikroskopie Glutaraldehyde 50 % solution in water for electron microscopy	Kat.Nr./Cat.No. : 23116 Lot/Contr.No. : 120662
--	---

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Summenformel formula		C ₅ H ₈ O ₂	
Molekulargewicht molecular weight		100.13	
Aussehen appearance		klare, farblose Lösung clear, colorless solution	entspricht corresponds
Gehalt (%) assay	Titration	45.0 – 53.0	50.7
Brechungsindex refractive index	20 °C	1.410 – 1.421	1.4164
Purification index	A235/ A280 1 % in H ₂ O		0.11
Mindesthaltbarkeit minimum shelf life			02/ 2015
Lagerung (°C) storage			+2 bis/ to +8

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 31.05.2012

Dipl.-Ing. (FH) Bernhard Göckel

Christian Monsler

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.