

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

| | | |
|------------------------------|--------------------------|---------------|
| Jodacetamid reinst | Kat.Nr./Cat.No. : | 26710 |
| Iodoacetamide research grade | Lot/Contr.No. : | 150498 |

| Parameter parameter | Methode method | Spezifikation specification | Ergebnis result |
|---|-------------------------|--|---------------------------|
| Summenformel formula | | C ₂ H ₄ INO | |
| Molekulargewicht molecular weight | | 185.0 | |
| Aussehen appearance | | weißes bis fast weißes, krist. Pulver oder Kristalle white to almost white, cryst. powder or crystals | entspricht corresponds |
| Gehalt (%) assay | HPLC | min. 99 | 100.0 |
| Trocknungsverlust (%) loss on drying | 50 °C, 4 h | max. 6 | 3.03 |
| Löslichkeit solubility | 1 % in H ₂ O | klar und farblos clear and colorless | entspricht corresponds |
| Identität identity | IR | entspricht corresponds | entspricht corresponds |
| Mindesthaltbarkeit minimum shelf life | | | 01/ 2018 |
| Lagerung (°C) storage | | | +2 bis/ to +8 |

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 24.04.2015

Dipl.-Ing. (FH) Bernhard Göckel

Christian Monsler

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.