

**ANALYSENZERTIFIKAT**  
CERTIFICATE OF ANALYSIS

<b>Kanadabalsam</b>	<b>Kat.Nr./Cat.No. :</b>	<b>26896</b>
Canada balsam for microscopy	<b>Lot/Contr.No. :</b>	<b>150914</b>

<b>Parameter</b> parameter	<b>Methode</b> method	<b>Spezifikation</b> specification	<b>Ergebnis</b> result
<b>Aussehen</b> appearance		klare, gelbe, viskose Flüssigkeit clear, yellow, viscous liquid	entspricht corresponds
<b>Dichte (g/ml)</b> density	20 °C	0.97 – 1.0	0.9917
<b>Brechungsindex</b> refractive index	20 °C	1.520 – 1.525	1.523
<b>Ursprung</b> origin		extrahiert aus den Rinden verschiedener Tannen extracted from the bark of various fir species (Abies canadensis, Abies balsamea)	
<b>Mindesthaltbarkeit</b> minimum shelf life			08/ 2020
<b>Lagerung (°C)</b> storage			+15 bis/ to +30

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.**  
**Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
We do not guarantee that the product can be used for a special application.  
This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date: 04.08.2015**

Dipl.-Ing. (FH) Bernhard Göckel

Daniela Lux-Helmstetter

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
This report has been computer generated and does not contain a signature.