

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

Kanadabalsam	Kat.Nr./Cat.No. : 26896
Canada balsam for microscopy	Lot/Contr.No. : 180012

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Aussehen appearance		klare, gelbe, viskose Flüssigkeit clear, yellow, viscous liquid	entspricht corresponds
Dichte (g/ml) density	20 °C	0.97 – 1.0	0.994
Brechungsindex refractive index	20 °C	1.520 – 1.525	1.523
Ursprung origin		extrahiert aus den Rinden verschiedener Tannen extracted from the bark of various fir species (Abies canadensis, Abies balsamea)	
Mindesthaltbarkeit minimum shelf life			04.01.2023
Lagerung (°C) storage			+15 bis/ to +30

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 22.01.2018

Dipl.-Ing. (FH) Bernhard Göckel

Christian Monsler

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.