

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

| | | |
|--|--------------------------|---------------|
| 3-Methacryloxypropyltrimethoxysilan | Kat.Nr./Cat.No. : | 28739 |
| 3-Methacryloxypropyltrimethoxysilane (Bind-Silane) | Lot/Contr.No. : | 120585 |

| Parameter parameter | Methode method | Spezifikation specification | Ergebnis result |
|---|-------------------|--|---------------------------|
| Summenformel formula | | C ₁₀ H ₂₀ O ₅ Si | |
| Molekulargewicht molecular weight | | 248.4 | |
| Aussehen appearance | | klare, farblose Flüssigkeit clear, colorless liquid | entspricht corresponds |
| Gehalt (%) assay | GC | 98.0 – 100.0 | 98.4 |
| Brechungsindex refractive index | 20 °C | 1.430 – 1.433 | 1.4312 |
| Farbzahl color | APHA | max. 30 | 15 |
| Viskosität (mm²/s) viscosity | | 2.3 – 2.5 | 2.4 |
| Mindesthaltbarkeit minimum shelf life | | | 05/ 2015 |
| Lagerung (°C) storage | | | +2 bis/ to +8 |

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 13.02.2014

Dipl.-Ing. (FH) Bernhard Göckel

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.