

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

Polyethylenimin 50% Lösung in Wasser pract.	Kat.Nr./Cat.No. :	33141
Polyethyleneimine 50% solution in water	Lot/Contr.No. :	201383

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Aussehen appearance		farblose bis blassgelbe viskose Flüssigkeit colorless to yellowish, viscous liquid	entspricht corresponds
pH	5 % (w/w) in H ₂ O	10 – 12	11
Nichtflüchtige Anteile (%) non-volatile matter	IR-Trocknung IR-drying	48.0 – 52.0	49.5
Scheinbare Viskosität (mPas) apparent viscosity		18 000 – 40 000	22 200
Mindesthaltbarkeit minimum shelf life			15.12.2023
Lagerung (°C) storage			+15 bis/ to +30

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 15.12.2020

Christian Monsler

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.