

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

Polyvinylpyrrolidon 30	Kat.Nr./Cat.No. : 33421
Polyvinylpyrrolidone 30	Lot/Contr.No. : 221020

Parameter parameter	Methode method	Spezifikation specification	Ergebnis result
Aussehen appearance		weißes bis gelbliches Pulver white to yellowish powder	entspricht corresponds
Identität identity	IR	entspricht corresponds	entspricht corresponds
K-Wert/ K-value		27.0 – 33.0	30.4
pH	10 % in H ₂ O	3.0 – 7.0	3.5
Mindesthaltbarkeit minimum shelf life			20.10.2027
Lagerung (°C) storage			+15 bis/ to +30

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
We do not guarantee that the product can be used for a special application.
This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control

Druckdatum/ printing date: 24.10.2022

Dr. Judith Koch

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
This report has been computer generated and does not contain a signature.