

A N A L Y S E N Z E R T I F I K A T
 CERTIFICATE OF ANALYSIS

| | | |
|---------------------------------|--------------------------|---------------|
| Propidiumjodid reinst | Kat.Nr./Cat.No. : | 33671 |
| Propidium iodide research grade | Lot/Contr.No. : | 201148 |

| Parameter parameter | Methode method | Spezifikation specification | Ergebnis result |
|---|-------------------|-----------------------------------|---------------------------|
| Summenformel formula | | $C_{27}H_{34}N_4^{2+} \cdot 2I^-$ | |
| Molekulargewicht molecular weight | | 668.40 | |
| Aussehen appearance | | rotes Pulver red powder | entspricht corresponds |
| Gehalt (%) assay | Titration | min. 90.0 | 95.36 |
| Mindesthaltbarkeit minimum shelf life | | | 23.10.2023 |
| Lagerung (°C) storage | | | +2 bis/ to +8 |

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.
 We do not guarantee that the product can be used for a special application.
 This document does not release you from performing the standard control upon receipt of incoming goods.

SERVA Electrophoresis GmbH
Qualitätskontrolle/ Quality Control
Druckdatum/ printing date: 27.10.2020

Christian Monsler

Daniela Lux-Helmstetter

Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.
 This report has been computer generated and does not contain a signature.