

**ANALYSENZERTIFIKAT**  
CERTIFICATE OF ANALYSIS

<b>Detergent 7 X<sup>®</sup> neutral, phosphatfrei</b>	<b>Kat.Nr./Cat.No. : 34205</b>
Detergent 7 X <sup>®</sup> neutral, phosphate-free	<b>Lot/Contr.No. : P120141</b>

<b>Parameter</b> parameter	<b>Methode</b> method	<b>Spezifikation</b> specification	<b>Ergebnis</b> result
<b>Aussehen</b> appearance		weiß-gelbliche, viskose Flüssigkeit white-yellowish, viscous liquid	entspricht corresponds
<b>Anwendung</b> use		Reinigung cleaning procedures	
<b>pH</b>	1:10 verd./ dil. in H <sub>2</sub> O	7.0 – 8.0	7.2
<b>Leitfähigkeit (mS/ cm)</b> conductivity	1:10 verd./ dil. in H <sub>2</sub> O	max. 5.0	1.5
<b>Mindesthaltbarkeit</b> minimum shelf life			03/ 2017
<b>Lagerung (°C)</b> storage			+15 bis/ to +30

**Farbunterschiede von lot zu lot sind möglich.**  
The colour may change from batch to batch.

**Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.**  
**Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.**  
We do not guarantee that the product can be used for a special application.  
This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date: 03.04.2012**

Dipl.-Ing. (FH) Bernhard Göckel

Christian Monsler

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
This report has been computer generated and does not contain a signature.