

**ANALYSENZERTIFIKAT**  
CERTIFICATE OF ANALYSIS

|                                                        |                                |
|--------------------------------------------------------|--------------------------------|
| <b>Detergent 7 X<sup>®</sup> neutral, phosphatfrei</b> | <b>Kat.Nr./Cat.No. : 34205</b> |
| Detergent 7 X <sup>®</sup> neutral, phosphate-free     | <b>Lot/Contr.No. : P160309</b> |

| Parameter<br>parameter                          | Methode<br>method                    | Spezifikation<br>specification                                               | Ergebnis<br>result        |
|-------------------------------------------------|--------------------------------------|------------------------------------------------------------------------------|---------------------------|
| <b>Aussehen</b><br>appearance                   |                                      | weiß-gelbliche,<br>viskose Flüssigkeit<br>white-yellowish,<br>viscous liquid | entspricht<br>corresponds |
| <b>Anwendung</b><br>use                         |                                      | Reinigung<br>cleaning procedures                                             |                           |
| <b>pH</b>                                       | 1:10 verd./ dil. in H <sub>2</sub> O | 7.0 – 8.0                                                                    | 7.2                       |
| <b>Leitfähigkeit (mS/ cm)</b><br>conductivity   | 1:10 verd./ dil. in H <sub>2</sub> O | max. 5.0                                                                     | 2.2                       |
| <b>Mindesthaltbarkeit</b><br>minimum shelf life |                                      |                                                                              | 28.07.2021                |
| <b>Lagerung (°C)</b><br>storage                 |                                      |                                                                              | +15 bis/ to +30           |

**Farbunterschiede von lot zu lot sind möglich.**  
The colour may change from batch to batch.

Die Eignung des Produktes für spezielle Anwendungszwecke wird nicht zugesichert.  
Dieses Dokument entbindet nicht von einer branchenüblichen Eingangskontrolle.  
We do not guarantee that the product can be used for a special application.  
This document does not release you from performing the standard control upon receipt of incoming goods.

**SERVA Electrophoresis GmbH**  
**Qualitätskontrolle/ Quality Control**

**Druckdatum/ printing date: 29.07.2016**

Dipl.-Ing. (FH) Bernhard Göckel

Patricia Milford

**Dieses Dokument ist ein Computerausdruck und deshalb ohne Unterschrift gültig.**  
This report has been computer generated and does not contain a signature.